



Cytomegalovirus PCR Detection Kit

For Research Use Only

Cat. No.: PK3001
Shipment: Wet Ice

Quantity: 50 Reactions
Storage: -20 °C

Description

This kit designed for qualitative detection of human sample by the method of polymerase chain reaction.

Kit Contents:

The kit for 50 amplification reactions consists of:

1. 1x PCR MIX	1000µl	4. Mineral Oil	2ml
2. Taq DNA polymerase	15µl (5u/µl)	5. Positive Control	50µl
3. DNase Free, Deionized Sterile Water	5ml		

The Kit should be stored at - 20°C.

DNA Extraction:

DNA can be extracted using DNP™ (# EX6071) or using other standard methods.

PCR Protocol:

1. Take out the kit and unfreeze the tubes, then put all the tubes on ice. Before opening tubes vortex and spin them. The final volume of each reaction will be 25µl.
2. Label new 0.5ml tubes for amplification reaction(s) for test(s), positive and negative control.
3. Add the following reagents for each tube on ice:

1x PCR MIX	19.7µl
Taq DNA polymerase	0.3µl

- Note: To avoid contamination all reagents must be taken with separate clean tips!

4. Mix the mixture thoroughly by shaking and spin.
5. To each tube add one-drop (20-25µl) **mineral oil**.
6. Add 5µl **DNA***(Use specified pipette for sampling of DNA).
7. Close tubes; spin the mixtures on microfuge for 3-5sec.
8. Transfer the tubes to preheated thermocycler and start the program:

Cycling parameters:

95°C - 180sec
62°C - 40sec
72°C - 40sec
1 cycle

93°C - 40sec
61°C - 40sec
72°C - 40sec
35 cycles

Analyze 6µl of amplified samples directly in a 2% agarose gel without adding loading buffer. The presence of 222bp fragments indicates positive test.

SinaClon BioScience

www.sinaclon.com

Tel: +98(0)21 4463 0050-51
+98(0)21 4466 5156
+98(0)2144666204

Mob.:+989023120052 ,54,56

Customer Service: +98(0)9023120059
+98(0)26336670907

Place your order: order@sinaclon.com